

**CERTIFICATE OF ANALYSIS**



**ANALYSIS PERFORMED FOR**

Aaron Nir  
Alytech  
100 Coman Street  
Rothwell QLD 4022

**CERTIFICATE NO.:** 386704

**ISSUE DATE:** 17/08/15

**REVISION NO:** 00

This certificate supercedes any previous revisions

**SAMPLE INFORMATION**

Description: Other - Waste  
Date Received: 17/08/2015  
Testing Commenced: 17/08/2015  
Order No:

**CONDITIONS OF SAMPLE ON RECEIPT**

Receipt Temperature: 25 °C (Surface Temperature taken by infra-red)  
Storage Temperature: 25 °C  
Sampled Date (if provided):

**RESULTS OF ANALYSIS**

*Sample(s) were analysed as received, and the results pertain only to the submitted sample(s).*

**Other - Waste**

CODE	Sample DESCRIPTION	Standard Plate Count (CFU/g)	Coliforms (CFU/g)	Escherichia coli (CFU/g)
386704-1	NT1	37,000,000	1,300,000	300,000
386704-2	NT2	74,000,000	7,000,000	3,000,000
386704-3	NT3	60,000,000	9,000,000	6,000,000
386704-4	ST1	60,000	1,500	900
386704-5	SC2	<10	<10	<10
386704-6	ALM3	<10	<10	<10
		M2.5	M8.8	M8.8
		AOAC 990.12	AOAC 991.14	AOAC 991.14

DEFINITIONS: > = Greater than < = Less than ~ = Estimated ND = Not detected MPN = Most probable number CFU = Colony forming units  
- = Not tested TBA = To be advised \* = Test not covered by scope of NATA accreditation RP = Result Pending, confirmatory testing in progress

NT1,2,3 - Not treated, ST1 - Sterimed, SC2 - Stericid, ALM3 - Alymed Steriliser

Hayden Smith, Microbiologist  
(Brisbane)



The results of the tests, calibrations and/or measurements included in this document are traceable to Australian/national standards.

Accredited for compliance with ISO/IEC 17025.

NATA Corporate Accreditation No.: 2455

HPC Holdings Pty Ltd trading as Symbio Alliance ABN 93 621 286 928

- Brisbane: 52 Brandl Street, Eight Mile Plains QLD 4113
- Melbourne: Unit 15, 640-680 Geelong Rd, Brooklyn VIC 3025
- Sydney: 1 Braidwood St, Enfield NSW 2136
- Rockhampton: Unit 3, 191 Berserker St, Berserker QLD 4701
- Tel: 1300 703 166
- Fax: +617 3219 0333